



Erin Viney
Mayer Environmental
Transport Avenue
Brentford
Middlesex
TW8 9HA

t: 020 8847 3637
f: 020 8847 3638
e: Mayer

i2 Analytical Ltd.
7 Woodshots Meadow,
Croxley Green
Business Park,
Watford,
Herts,
WD18 8YS

t: 01923 225404
f: 01923 237404
e: reception@i2analytical.com

Analytical Report Number : 19-72381

Project / Site name:	EMR Boreham	Samples received on:	18/11/2019
Your job number:	73580.004	Samples instructed on:	18/11/2019
Your order number:	018110	Analysis completed by:	26/11/2019
Report Issue Number:	1	Report issued on:	26/11/2019
Samples Analysed:	3 water samples		

Signed: *Karolina Marek*

Karolina Marek
Technical Reviewer (Reporting Team)

For & on behalf of i2 Analytical Ltd.

Standard Geotechnical, Asbestos and Chemical Testing Laboratory located at: ul. Pionierów 39, 41 -711 Ruda Śląska, Poland.

Accredited tests are defined within the report, opinions and interpretations expressed herein are outside the scope of accreditation.

Standard sample disposal times, unless otherwise agreed with the laboratory, are :

soils	- 4 weeks from reporting
leachates	- 2 weeks from reporting
waters	- 2 weeks from reporting
asbestos	- 6 months from reporting

Excel copies of reports are only valid when accompanied by this PDF certificate.

Any assessments of compliance with specifications are based on actual analytical results with no contribution from uncertainty of measurement. Application of uncertainty of measurement would provide a range within which the true result lies. An estimate of measurement uncertainty can be provided on request.

Iss No 19-72381-1 EMR Boreham 73580.004

This certificate should not be reproduced, except in full, without the express permission of the laboratory.

The results included within the report are representative of the samples submitted for analysis.

Page 1 of 4



Analytical Report Number: 19-72381

Project / Site name: EMR Boreham

Your Order No: 018110

Lab Sample Number				1364822	1364823	1364824		
Sample Reference				DP1	UPST1	DNST1		
Sample Number				None Supplied	None Supplied	None Supplied		
Depth (m)				None Supplied	None Supplied	None Supplied		
Date Sampled				18/11/2019	18/11/2019	18/11/2019		
Time Taken				None Supplied	None Supplied	None Supplied		
Analytical Parameter (Water Analysis)	Units	Limit of detection	Accreditation Status					

General Inorganics

pH	pH Units	N/A	ISO 17025	7.3	7.9	7.9		
Electrical Conductivity at 20 °C	µS/cm	10	ISO 17025	1000	740	650		
Total Cyanide (Low Level)	mg/l	0.001	ISO 17025	0.002	< 0.001	< 0.001		
Sulphate as SO ₄	mg/l	0.045	ISO 17025	191	86.0	84.8		
Sulphide	mg/l	0.005	NONE	0.43	< 0.005	< 0.005		
Chloride	mg/l	0.15	ISO 17025	100	80	87		
Total Phosphate as P	mg/l	0.02	ISO 17025	< 0.02	< 0.02	0.02		
Ammoniacal Nitrogen as N	mg/l	0.015	ISO 17025	2.9	0.024	0.030		
Ammonia as NH ₃	mg/l	0.015	ISO 17025	3.479	0.029	0.036		
Nitrate as N	mg/l	0.01	ISO 17025	0.25	2.08	2.30		
Nitrate as NO ₃	mg/l	0.05	ISO 17025	1.12	9.21	10.2		
Alkalinity	mgCaCO ₃ /l	3	ISO 17025	170	170	190		
Chemical Oxygen Demand (Total)	mg/l	2	ISO 17025	150	19	18		
BOD (Biochemical Oxygen Demand) (Total) - PL	mg/l	1	ISO 17025	150	1.5	1.4		
Total Suspended Solids	mg/l	2	ISO 17025	< 2.0	30	< 2.0		

Total Phenols

Total Phenols (monohydric)	mg/l	0.01	ISO 17025	< 0.010	< 0.010	< 0.010		
----------------------------	------	------	-----------	---------	---------	---------	--	--

Heavy Metals / Metalloids

Arsenic (dissolved)	mg/l	0.00015	ISO 17025	0.0007	0.0010	0.0011		
Boron (dissolved)	mg/l	0.01	ISO 17025	0.91	0.02	0.02		
Cadmium (dissolved)	mg/l	0.00002	ISO 17025	0.0001	0.00002	< 0.00002		
Chromium (dissolved)	mg/l	0.0002	ISO 17025	0.0016	0.0017	0.0011		
Copper (dissolved)	mg/l	0.0005	ISO 17025	0.0012	0.0017	0.0026		
Iron (dissolved)	mg/l	0.004	ISO 17025	0.35	0.090	0.095		
Lead (dissolved)	mg/l	0.0002	ISO 17025	0.0058	0.0002	0.0010		
Manganese (dissolved)	mg/l	0.00005	ISO 17025	0.12	0.037	0.036		
Mercury (dissolved)	mg/l	0.00005	ISO 17025	0.0003	< 0.00005	< 0.00005		
Nickel (dissolved)	mg/l	0.0005	ISO 17025	0.0085	0.0044	0.0042		
Selenium (dissolved)	mg/l	0.0006	ISO 17025	0.0021	0.0008	0.0009		
Zinc (dissolved)	mg/l	0.0005	ISO 17025	0.0167	0.0015	0.0026		

Petroleum Hydrocarbons

Mineral Oil (C10 - C40)	mg/l	0.01	NONE	4.31	< 0.01	< 0.01		
-------------------------	------	------	------	------	--------	--------	--	--

U/S = Unsuitable Sample I/S = Insufficient Sample



Analytical Report Number : 19-72381

Project / Site name: EMR Boreham

Water matrix abbreviations: Surface Water (SW) Potable Water (PW) Ground Water (GW) Process Water (PrW)

Analytical Test Name	Analytical Method Description	Analytical Method Reference	Method number	Wet / Dry Analysis	Accreditation Status
Alkalinity in Water (by discreet analyser)	Determination of Alkalinity by discreet analyser (colorimetry). Accredited matrices: SW, PW, GW.	In house method based on MEWAM & USEPA Method 310.2.	L082-PL	W	ISO 17025
Ammonia as NH ₃ in water	Determination of Ammonium/Ammonia/ Ammoniacal Nitrogen by the colorimetric salicylate/nitroprusside method. Accredited matrices SW, GW, PW.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L082-PL	W	ISO 17025
Ammoniacal Nitrogen as N in water	Determination of Ammonium/Ammonia/ Ammoniacal Nitrogen by the discrete analyser (colorimetric) salicylate/nitroprusside method. Accredited matrices SW, GW, PW.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L082-PL	W	ISO 17025
Biological oxygen demand (total) of water	Determination of biochemical oxygen demand in water (5 days). Accredited matrices: SW, PW, GW.	In-house method based on standard method 5210B.	L086-PL	W	ISO 17025
Boron in water	Determination of boron in water by acidification followed by ICP-OES. Accredited matrices: SW PW GW	In-house method based on MEWAM	L039-PL	W	ISO 17025
Chemical Oxygen Demand in Water (Total)	Determination of total COD in water by reflux oxidation with acidified K ₂ Cr ₂ O ₇ followed by colorimetry. Accredited matrices: SW, PW, GW.	HACH DR/890 Colorimeter Procedures Manual (48470-22) (Ref 0170.2)	L065-PL	W	ISO 17025
Chloride in water	Determination of Chloride colorimetrically by discrete analyser.	In house based on MEWAM Method ISBN 0117516260. Accredited matrices: SW, PW, GW.	L082-PL	W	ISO 17025
Electrical conductivity at 20oC of water	Determination of electrical conductivity in water by electrometric measurement. Accredited Matrices SW, GW, PW	In-house method	L031-PL	W	ISO 17025
EPH C8-C40 (with Min. Oil by calc.) water	Determination of dichloromethane/hexane extractable hydrocarbons in soil by GC-MS.	In-house method	L070-PL	W	NONE
Low level total cyanide in water	Determination of total cyanide by distillation followed by colorimetry. Accredited matrices: SW PW GW	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton (Skalar)	L080-PL	W	ISO 17025
Metals in water by ICP-MS (dissolved)	Determination of metals in water by acidification followed by ICP-MS. Accredited Matrices: SW, GW, PW except B=SW,GW, Hg=SW,PW, Al=SW,PW.	In-house method based on USEPA Method 6020 & 200.8 "for the determination of trace elements in water by ICP-MS.	L012-PL	W	ISO 17025
Metals in water by ICP-OES (dissolved)	Determination of metals in water by acidification followed by ICP-OES. Accredited Matrices SW, GW, PW, PrW.(Al, Cu,Fe,Zn).	In-house method based on MEWAM 2006 Methods for the Determination of Metals in Soil.	L039-PL	W	ISO 17025
Monohydric phenols in water	Determination of phenols in water by continuous flow analyser. Accredited matrices: SW PW GW	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton (skalar)	L080-PL	W	ISO 17025
Nitrate as N in water	Determination of nitrate by reaction with sodium salicylate and colorimetry. Accredited matrices SW, GW, PW.	In-house method based on Examination of Water and Wastewater & Polish Standard Method PN-82/C-04579.08,	L078-PL	W	ISO 17025
Nitrate in water	Determination of nitrate by reaction with sodium salicylate and colorimetry. Accredited matrices SW, GW, PW	In-house method based on Examination of Water and Wastewater & Polish Standard Method PN-82/C-04579.08,	L078-PL	W	ISO 17025
pH at 20oC in water (automated)	Determination of pH in water by electrometric measurement. Accredited matrices: SW PW GW	In-house method based on BS1377 Part 3, 1990, Chemical and Electrochemical Tests	L099-PL	W	ISO 17025
Sulphate in water	Determination of sulphate in water by acidification followed by ICP-OES. Accredited matrices: SW PW GW, PrW.	In-house method based on MEWAM 2006 Methods for the Determination of Metals in Soil.	L039-PL	W	ISO 17025

Iss No 19-72381-1 EMR Boreham 73580.004

This certificate should not be reproduced, except in full, without the express permission of the laboratory.

The results included within the report are representative of the samples submitted for analysis.

Page 3 of 4



Analytical Report Number : 19-72381

Project / Site name: EMR Boreham

Water matrix abbreviations: Surface Water (SW) Potable Water (PW) Ground Water (GW) Process Water (PrW)

Analytical Test Name	Analytical Method Description	Analytical Method Reference	Method number	Wet / Dry Analysis	Accreditation Status
Sulphide in water in mg l	Determination of sulphide in water by ion selective electrode.	In-house method	L029-PL	W	NONE
Suspended solids in water	Determined gravimetrically with GFC filtration papers.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L004-PL	W	ISO 17025
Total Phosphate as P in water	Determination of ortho phosphate in water by addition of ammonium molybdate, potassium antimonyl tartrate and ascorbic acid followed by colorimetry. Accredited matrices: SW, PW, GW.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L082-PL	W	ISO 17025

For method numbers ending in 'UK' analysis have been carried out in our laboratory in the United Kingdom.

For method numbers ending in 'PL' analysis have been carried out in our laboratory in Poland.

Soil analytical results are expressed on a dry weight basis. Where analysis is carried out on as-received the results obtained are multiplied by a moisture correction factor that is determined gravimetrically using the moisture content which is carried out at a maximum of 30oC.