



**Rebecca Beddard**  
Mayer Environmental  
Transport Avenue  
Brentford  
Middlesex  
TW8 9HA

**t:** 020 8847 3637  
**f:** 020 8847 3638  
**e:** Mayer

i2 Analytical Ltd.  
7 Woodshots Meadow,  
Croxley Green  
Business Park,  
Watford,  
Herts,  
WD18 8YS

**t:** 01923 225404  
**f:** 01923 237404  
**e:** reception@i2analytical.com

## **Analytical Report Number : 19-62460**

<b>Project / Site name:</b>	EMR Boreham	<b>Samples received on:</b>	25/09/2019
<b>Your job number:</b>	73580.003	<b>Samples instructed on:</b>	25/09/2019
<b>Your order number:</b>	018007	<b>Analysis completed by:</b>	04/10/2019
<b>Report Issue Number:</b>	1	<b>Report issued on:</b>	04/10/2019
<b>Samples Analysed:</b>	3 water samples		

**Signed:** *Karolina Marek*

Karolina Marek  
Technical Reviewer (Reporting Team)  
**For & on behalf of i2 Analytical Ltd.**

Standard Geotechnical, Asbestos and Chemical Testing Laboratory located at: ul. Pionierów 39, 41 -711 Ruda Śląska, Poland.

Accredited tests are defined within the report, opinions and interpretations expressed herein are outside the scope of accreditation.

Standard sample disposal times, unless otherwise agreed with the laboratory, are :

soils	- 4 weeks from reporting
leachates	- 2 weeks from reporting
waters	- 2 weeks from reporting
asbestos	- 6 months from reporting

Excel copies of reports are only valid when accompanied by this PDF certificate.

Any assessments of compliance with specifications are based on actual analytical results with no contribution from uncertainty of measurement. Application of uncertainty of measurement would provide a range within which the true result lies. An estimate of measurement uncertainty can be provided on request.



Analytical Report Number: 19-62460

Project / Site name: EMR Boreham

Your Order No: 018007

Lab Sample Number	13593	13594	13595		
Sample Reference	DP1	UPST1	ONST1		
Sample Number	None Supplied	None Supplied	None Supplied		
Depth (m)	None Supplied	None Supplied	None Supplied		
Date Sampled	25/09/2019	25/09/2019	25/09/2019		
Time Taken	None Supplied	None Supplied	None Supplied		
Analytical Parameter (Water Analysis)	Units	Limit of detection	Accreditation Status		
<b>WATERS</b>					

**General Inorganics**

	pH Units	N/A	ISO 17025	7.9	8.3	8.1		
pH								
Electrical Conductivity at 20 °C	µS/cm	10	ISO 17025	820	610	620		
Total Cyanide (Low Level)	mg/l	0.001	ISO 17025	< 0.001	< 0.001	< 0.001		
Sulphate as SO <sub>4</sub>	mg/l	0.045	ISO 17025	133	118	57.4		
Sulphide	mg/l	0.005	NONE	< 0.005	< 0.005	< 0.005		
Chloride	mg/l	0.15	ISO 17025	100	78	78		
Total Phosphate as P	mg/l	0.02	ISO 17025	0.05	0.06	0.07		
Ammoniacal Nitrogen as N	mg/l	0.015	ISO 17025	3.0	0.079	0.080		
Ammonia as NH <sub>3</sub>	mg/l	0.015	ISO 17025	3.630	0.096	0.098		
Nitrate as N	mg/l	0.01	ISO 17025	0.19	1.71	1.59		
Nitrate as NO <sub>3</sub>	mg/l	0.05	ISO 17025	0.86	7.59	7.06		
Alkalinity	mgCaCO <sub>3</sub> /l	3	ISO 17025	150	180	180		
Chemical Oxygen Demand (Total)	mg/l	2	ISO 17025	290	17	14		
BOD (Biochemical Oxygen Demand) (Total) - PL	mg/l	1	ISO 17025	210	2.4	2.2		
Total Suspended Solids	mg/l	2	ISO 17025	150	I/S	I/S		

**Total Phenols**

Total Phenols (monohydric)	mg/l	0.01	ISO 17025	< 0.010	< 0.010	< 0.010		
----------------------------	------	------	-----------	---------	---------	---------	--	--

**Heavy Metals / Metalloids**

Iron (total)	mg/l	0.004	ISO 17025	3.4	0.28	0.27		
Arsenic (total)	mg/l	0.00015	ISO 17025	0.0024	0.0029	0.0030		
Boron (total)	mg/l	0.01	ISO 17025	1.1	0.05	0.06		
Cadmium (total)	mg/l	0.00002	ISO 17025	0.00198	< 0.00002	0.00002		
Chromium (total)	mg/l	0.0002	ISO 17025	0.0122	0.0004	0.0004		
Copper (total)	mg/l	0.0005	ISO 17025	0.048	0.016	0.027		
Mercury (total)	mg/l	0.00005	ISO 17025	0.00087	< 0.00005	< 0.00005		
Manganese (total)	mg/l	0.00005	ISO 17025	0.21	0.031	0.051		
Nickel (total)	mg/l	0.0005	ISO 17025	0.0192	0.0032	0.0049		
Lead (total)	mg/l	0.001	ISO 17025	0.20	< 0.001	0.002		
Selenium (total)	mg/l	0.004	ISO 17025	< 0.004	< 0.004	< 0.004		
Zinc (total)	mg/l	0.0005	ISO 17025	1.00	0.0059	0.0145		

**Petroleum Hydrocarbons**

Mineral Oil (C10 - C40)	mg/l	0.01	NONE	< 0.01	< 0.01	< 0.01		
-------------------------	------	------	------	--------	--------	--------	--	--

U/S = Unsuitable Sample I/S = Insufficient Sample



**Analytical Report Number : 19-62460**

**Project / Site name: EMR Boreham**

**Water matrix abbreviations: Surface Water (SW) Potable Water (PW) Ground Water (GW) Process Water (PrW)**

Analytical Test Name	Analytical Method Description	Analytical Method Reference	Method number	Wet / Dry Analysis	Accreditation Status
Alkalinity in Water	Determination of Alkalinity by discrete analyser (colorimetry). Accredited matrices: SW, PW, GW.	In house method based on MEWAM & USEPA Method 310.2.	L082-PL	W	ISO 17025
Ammonia as NH3 in water	Determination of Ammonium/Ammonia/Ammoniacal Nitrogen by the colorimetric salicylate/nitroprusside method. Accredited matrices SW, GW, PW.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L082-PL	W	ISO 17025
Ammoniacal Nitrogen as N in water	Determination of Ammonium/Ammonia/Ammoniacal Nitrogen by the discrete analyser (colorimetric) salicylate/nitroprusside method. Accredited matrices SW, GW, PW.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L082-PL	W	ISO 17025
Biological oxygen demand (total) of water	Determination of biochemical oxygen demand in water (5 days). Accredited matrices: SW, PW, GW.	In-house method based on standard method 5210B. Samples received > 24 hrs after sampling, data may not be valid and should be interpreted with care.	L086-PL	W	ISO 17025
Chemical Oxygen Demand in Water (Total)	Determination of total COD in water by reflux oxidation with acidified K2Cr2O7 followed by colorimetry. Accredited matrices: SW, PW, GW.	HACH DR/890 Colorimeter Procedures Manual (48470-22) (Ref 0170.2)	L065-PL	W	ISO 17025
Chloride in water	Determination of Chloride colorimetrically by discrete analyser.	In house based on MEWAM Method ISBN 0117516260. Accredited matrices: SW, PW, GW.	L082-PL	W	ISO 17025
Electrical conductivity at 20oC of water	Determination of electrical conductivity in water by electrometric measurement. Accredited Matrices SW, GW, PW	In-house method	L031-PL	W	ISO 17025
EPH C8-C40 (with Min. Oil by calc.) water	Determination of dichloromethane/hexane extractable hydrocarbons in soil by GC-MS.	In-house method	L070-PL	W	NONE
Low level total cyanide in water	Determination of total cyanide by distillation followed by colorimetry. Accredited matrices: SW PW GW	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton (Skalar)	L080-PL	W	ISO 17025
Metals in water by ICP-MS	Determination of metals in water by acidification followed by ICP-MS. Accredited Matrices SW, GW, PW.	In-house method based on USEPA Method 6020 & 200.8 "for the determination of trace elements in water by ICP-MS"	L012-UK	W	ISO 17025
Metals in water by ICP-MS (total)	Determination of metals in water by acidification followed by ICP-MS. Accredited Matrices: SW, GW, PW except B=SW,GW, Hg=SW,PW, Al=SW,PW.	In-house method based on USEPA Method 6020 & 200.8 "for the determination of trace elements in water by ICP-MS.	L012-PL	W	ISO 17025
Metals in water by ICP-OES (total)	Determination of metals in water by acidification followed by ICP-OES. Accredited matrices: SW PW GW, PrW (Al, Fe, Cu, Zn).	In-house method based on MEWAM 2006 Methods for the Determination of Metals in Soil.	L039-PL	W	ISO 17025
Monohydric phenols in water	Determination of phenols in water by continuous flow analyser. Accredited matrices: SW PW GW	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton (skalar)	L080-PL	W	ISO 17025
Nitrate as N in water	Determination of nitrate by reaction with sodium salicylate and colorimetry. Accredited matrices SW, GW, PW.	In-house method based on Examination of Water and Wastewater & Polish Standard Method PN-82/C-04579.08,	L078-PL	W	ISO 17025
Nitrate in water	Determination of nitrate by reaction with sodium salicylate and colorimetry. Accredited matrices SW, GW, PW	In-house method based on Examination of Water and Wastewater & Polish Standard Method PN-82/C-04579.08,	L078-PL	W	ISO 17025
pH at 20oC in water (automated)	Determination of pH in water by electrometric measurement. Accredited matrices: SW PW GW	In-house method based on BS1377 Part 3, 1990, Chemical and Electrochemical Tests	L099-PL	W	ISO 17025
Sulphate in water	Determination of sulphate in water by acidification followed by ICP-OES. Accredited matrices: SW PW GW, PrW.	In-house method based on MEWAM 2006 Methods for the Determination of Metals in Soil.	L039-PL	W	ISO 17025



**Analytical Report Number : 19-62460**

**Project / Site name: EMR Boreham**

**Water matrix abbreviations: Surface Water (SW) Potable Water (PW) Ground Water (GW) Process Water (PrW)**

Analytical Test Name	Analytical Method Description	Analytical Method Reference	Method number	Wet / Dry Analysis	Accreditation Status
Sulphide in water in mg l	Determination of sulphide in water by ion selective electrode.	In-house method	L029-PL	W	NONE
Suspended solids in water	Determined gravimetrically with GFC filtration papers.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L004-PL	W	ISO 17025
Total Phosphate as P in water	Determination of ortho phosphate in water by addition of ammonium molybdate, potassium antimonyl tartrate and ascorbic acid followed by colorimetry. Accredited matrices: SW, PW, GW.	In-house method based on Examination of Water and Wastewater 20th Edition: Clesceri, Greenberg & Eaton	L082-PL	W	ISO 17025

**For method numbers ending in 'UK' analysis have been carried out in our laboratory in the United Kingdom.**

**For method numbers ending in 'PL' analysis have been carried out in our laboratory in Poland.**

**Soil analytical results are expressed on a dry weight basis. Where analysis is carried out on as-received the results obtained are multiplied by a moisture correction factor that is determined gravimetrically using the moisture content which is carried out at a maximum of 30°C.**