



**Title: Monitoring of Particulate Matter (PM)  
& Gaseous Emissions**

**Permit Number:** TBC  
**Operator:** CBE Plus Limited  
**Installation:** Spray Booth  
**Monitoring Dates:** 11 November 2025

Reference Number: EI/10720

Client Organisation: CBE Plus Limited  
Address: Enterprise Drive  
Holmewood  
Chesterfield  
S42 5UZ

Monitoring Organisation: CES Environmental Instruments Ltd  
Address: Bretby Business Park  
Ashby Road  
Burton on Trent  
Staffordshire  
DE15 0YZ

Date of Report: 1 December 2025

Report Prepared By: Martin Rodgers  
MCERTS Registration Number: MM 04 531 (Level 2, TE1, TE2, TE3, TE4)

Signed: .....

Report Approved By: Robert Allen  
MCERTS Registration Number: MM 02 009 (Level 2, TE1, TE2, TE3, TE4)

Signed: .....

**Contents**

Part 1: Executive Summary .....3  
1.1 Monitoring Objectives .....3  
1.2 Monitoring Results .....4  
1.3 Operating Information .....5  
1.4 Monitoring Deviations.....5

Part 2: Supporting Information.....6  
Appendix 1 General Information .....6  
Appendix 2 Diagrams of Emission Point .....8  
Appendix 3 Particulate Matter (Sampling Measurement & Results) .....10  
Appendix 4 Gas (Instrument Measured Values & Results).....25  
Appendix 5 Instrumental Gas Analyser Site Calibration Data .....28  
Appendix 6 Calibration Certificates .....31  
Appendix 7 Uncertainty Calculations.....36

## **Part 1: Executive Summary**

### **1.1 Monitoring Objectives**

CBE Plus Limited placed a contract with CES Environmental Instruments Ltd for the compliance check monitoring of emissions to air from the Stripping Line.

#### **Stripping Line**

##### **Plating Process**

CBE+ is a global leader in precision Electroless Nickel Plating (ENP), complemented by precision CNC machining and a range of specialist ancillary services.

The test work was undertaken on 11 November 2025 by CES Environmental Instruments Ltd Engineers and carried out as part of CES Environmental Instruments Ltd job reference EI/10720.

The substances monitored were: -

Particulate Matter

Velocity Volume Flow

Total Organic Compounds

On the day of testing there were no special requirements for the monitoring.

## 1.2 Monitoring Results

Emission Point Reference: Spray Booth (Left Hand Side)

Substance to be Monitored	Emission Limit Value	Periodic Monitoring Result	Uncertainty of Measurement (95% CI)	Blank Result	Units	Reference Conditions	Emission Rate	Date of Sampling	Start and End Times	Monitoring Method Reference	Accreditation for use of Method	Operating Status
Particulate Matter	To Be Confirmed	0.09	0.13	0.03*	mg/m <sup>3</sup>	273K, 101.3kPa	0.001kg/hr	11 November 2025	09:48-12:57	BS EN 13284-1	UKAS & MCERTS	Normal Operation
Velocity	Not Applicable	5.53	0.40		m/s	Actual	-	11 November 2025	09:48-12:57	BS EN 16911	UKAS & MCERTS	Normal Operation
Mean Volume Flow	Not Applicable	7692.3	-	-	m <sup>3</sup> /hr	Actual	-	11 November 2025	09:48-12:57	BS EN 16911	UKAS & MCERTS	Normal Operation
Total Organic Compound	To Be Confirmed	114.69	8.95	-	mg/m <sup>3</sup>	273K, 101.3kPa	-	11 November 2025	10:20-11:19	BS EN 12619	UKAS & MCERTS	Normal Operation

\* Indicates where a value less than the limit of detection of the weighing procedure (0.09mg) has been reported, the value lies between the detection limit and zero. A value of half the limit of detection (0.045mg) has been used to calculate the concentration.

Emission Point Reference: Spray Booth (Right Hand Side)

Substance to be Monitored	Emission Limit Value	Periodic Monitoring Result	Uncertainty of Measurement (95% CI)	Blank Result	Units	Reference Conditions	Emission Rate	Date of Sampling	Start and End Times	Monitoring Method Reference	Accreditation for use of Method	Operating Status
Particulate Matter	To Be Confirmed	0.38	0.15	0.04*	mg/m <sup>3</sup>	273K, 101.3kPa	0.004kg/hr	11 November 2025	09:50-13:02	BS EN 13284-1	UKAS & MCERTS	Normal Operation
Velocity	Not Applicable	7.80	0.34		m/s	Actual	-	11 November 2025	09:50-13:02	BS EN 16911	UKAS & MCERTS	Normal Operation
Mean Volume Flow	Not Applicable	10810.7	-	-	m <sup>3</sup> /hr	Actual	-	11 November 2025	09:50-13:02	BS EN 16911	UKAS & MCERTS	Normal Operation
Total Organic Compound	To Be Confirmed	64.33	8.34	-	mg/m <sup>3</sup>	273K, 101.3kPa	-	11 November 2025	11:35-12:34	BS EN 12619	UKAS & MCERTS	Normal Operation

\* Indicates where a value less than the limit of detection of the weighing procedure (0.09mg) has been reported, the value lies between the detection limit and zero. A value of half the limit of detection (0.045mg) has been used to calculate the concentration.

### 1.3 Operating Information

Emission Point Reference: Spray Booth (Left Hand / Right Hand)

Process Type	Batch Sample Details	Fuel	Product	Load	Abatement
Continuous	-	-	Metal Components	Normal Operation	None

Comparison of Operator Continuous Emissions Monitoring Systems (CEMS) and Periodic Monitoring Results								
Substance to be Monitored	Emission Limit Value	Periodic Monitoring Result	Uncertainty of Measurement (95% CI)	Units	Reference Conditions	Date of Sampling	Start and End Times	CEMS Results
No CEMS Installed								

### 1.4 Monitoring Deviations

The sample plane does not comply upstream and downstream as per the requirements of BS EN 15259.

## **Part 2: Supporting Information**

### **Appendix 1 General Information**

### CES Environmental Instruments Ltd staff details

Name	Role	MCERT Registration Number	Level 1	Level 2	TE1	TE2	TE3	TE4	At site
Paul Broxham	Team Leader	MM 13 1254		✓	✓	✓	✓	✓	T
				Feb 2030	Feb 2030	Feb 2030	Nov 2029	Nov 2029	
Dylan Foster	Technician	Trainee							✓

T = Nominated Team Leader on Site

### CES Environmental Instruments Ltd method details

Pollutant	Method	CES Procedure
Particulate Matter	BS EN 13284-1	WI 4/1
Velocity Volume Flow	BS EN 16911-1	WI4/56
Total Organic Compound	BS EN 12619	WI4/28

### Monitoring Equipment Used

Gravimat & Probe  
Sick Maihak FID

CES Environmental Instruments Ltd Reference: C300 & C193  
CES Environmental Instruments Ltd Reference: C301

## **Appendix 2**

### **Diagrams of Emission Point**

## Sampling Location

Dimensions	Cross Sectional Area	Orientation	Sample Ports Available/Used	Sampling Positions Per Plane	Standard
Dia = 700mm	0.385m <sup>2</sup>	Vertical	PM-2/2 Gases-2/1	4 1	BS EN 15259
<p><b>Comments:</b> Sample ports: 2 off 4" BSP sockets</p> <p>Sample times are calculated from the total sample time equally divided by the no. of sample positions per plane. The minimum sample time per position is 3 minutes.</p> <p>Sample positions calculated using the Tangential method for circular ducts.</p> <p><b>Pitot Traverse</b> Along lines A &amp; B at positions consistent with BS EN 15259 these positions are: 14.6%, 85.4%</p> <p><b>Sample Positions</b> Along lines A &amp; B at as many of the positions required within the standard method as can be achieved given the clearance limitations behind each socket. BS EN 15259 requires sampling at 4 points (2 on each of two lines) these positions are: 14.6%, 85.4%</p>					
				Yes	No
<b>Has homogeneity test been carried out?</b>					✓
<b>If Yes - Is stack gas homogenous?</b>					
<b>Any physical or regulatory restrictions regarding usage of equipment?</b> N/A					

Compliance with BS EN 15259 / EA TGN M1	Yes	No
<b>Does the sample plane comply upstream?</b>		✓*
<b>Does the sample plane comply downstream?</b>		✓*
<b>Are the appropriate sample ports fitted?</b>	✓	
<b>Do the stack gas velocity / temperature profiles comply?</b>	✓	
<b>Minimum platform area &gt;5m<sup>2</sup></b>	✓	

\*BS 15259 has a general requirement that the sampling plane shall be in a straight section of duct at least 5 hydraulic diameters downstream and 2 hydraulic diameters upstream of any bend or obstruction which could produce turbulent flow.



**Appendix 3**  
**Particulate Matter (Sampling Measurement & Results)**

**Site :** CBE Plus Limited  
**Date :** 11 November 2025  
**Plant :** Spray Booth Left Hand Side  
**File Ref.** 10720

**Mean Particulate Results**

Filter	Time	mg/m <sup>3</sup> (Actual Conditions)	m <sup>3</sup> /hr	mg/Nm <sup>3</sup> (Reference Conditions)	Nm <sup>3</sup> /hr	kg/hr
100338	09:48-10:50	0.04	8004	0.05	7059.0	0.000
100132	10:52-11:54	0.10	7935	0.11	6998.0	0.001
1001242	11:55-12:57	0.10	7138	0.11	6295.0	0.001
	<b>Mean</b>	0.08	7692.3	0.09	6784.0	0.001

**Control Blank Filter**

Filter	Volume (m <sup>3</sup> )	
100338	1.448	
100132	1.423	
1001242	1.284	
<b>Mean</b>	<b>1.385</b>	(Reference Conditions with no correction for Oxygen)

**Filter** 1152051  
  
**Tare Weight** 18264.80 mg  
**Gross Weight** 18264.85 mg \*  
  
**Gain** 0.04 mg  
  
**Measured Oxygen** %  
  
**Concentration** 0.03 mg/Nm<sup>3</sup>\*

Results Correct to

Temperature		Pressure		Oxygen		Gas	
°C/K	0/273	mbar/kPa	1013/101.3	%		Wet/Dry	Wet

\* Indicates where a value less than the limit of detection of the weighing procedure (0.09mg) has been reported, the value lies somewhere between the detection limit and zero. A value of half the limit of detection (0.045mg) has been used to calculate the concentration.

**Velocity Results**

	Time	m/s
<b>Test 1</b>	09:48-10:50	5.80
<b>Test 2</b>	10:52-11:54	5.70
<b>Test 3</b>	11:55-12:57	5.10
	<b>Mean</b>	<b>5.53</b>

protocol simultaneous isokinetic extraction measurement

11/11/2025 09:48

11/11/2025 10:50

collector-no. 338  
 engineer PB DF  
 plant name Spray Boot LS  
 place CBE Plus  
 remarks Normal Operation  
 Dia = 700mm  
 T1

operating parameter

normal density humid [ kg / m<sup>3</sup> ] : 1.27  
 water vapour [ %Vol ] : 2.32  
 ambient pressure [ mbar ] : 963  
 duct cross-section [ m<sup>2</sup> ] : 0.385

extraction parameter

change of meas. point [h:m:s] : 00:15:00  
 points / axis : 2  
 nozzles diameter [ mm ] : 10  
 isokinetic factor : 1  
 tare weight [ mg ] : 17774.53  
 gross weight [ mg ] : 17774.6

evaluation

meas. time [h:m:s] : 01:00:00  
 dust weight [ mg ] : 0.07

extracted partial volume  
 actual conditions [ m<sup>3</sup> ] : 1.642  
 in norm wet [ Nm<sup>3</sup> ] : 1.448  
 in norm dry [ Nm<sup>3</sup> ] : 1.414

volume flow in duct  
 actual conditions [m<sup>3</sup>/h] : 8004  
 in norm wet [Nm<sup>3</sup>/h] : 7059  
 in norm dry [Nm<sup>3</sup>/h] : 6895

dust concentration  
 actual conditions [mg/m<sup>3</sup>] : 0.04  
 in norm wet [mg/Nm<sup>3</sup>] : 0.05  
 in norm dry [mg/Nm<sup>3</sup>] : 0.05

protocol simultaneous isokinetic extraction measurement

11/11/2025 09:48

11/11/2025 10:50

measured values table

axis	depth	T_probe [°C]	v_duct [m/s]	angle [grd]	Q_act. [m³/h]	volume [m³]	meas. time [ H:M:S ]	p10 [mbar]	p40 [mbar]
1	1	22	5.8	0.1	1.64	0.412	00:15:00	1.00	-315
1	2	22	5.7	0.1	1.63	0.406	00:15:00	1.00	-311
2	1	22	5.8	0.9	1.64	0.412	00:15:00	1.00	-315
2	2	22	5.8	0.2	1.65	0.412	00:15:00	1.00	-319
		22	5.8	0.3	1.64	0.410		1.00	-315

protocol simultaneous isokinetic extraction measurement

11/11/2025 10:52

11/11/2025 11:54

collector-no. 132  
 engineer PB DF  
 plant name Spray Boot LS  
 place CBE Plus  
 remarks Normal Operation  
 Dia = 700mm  
 T2

operating parameter

normal density humid [ kg / m<sup>3</sup> ] : 1.27  
 water vapour [ %Vol ] : 2.32  
 ambient pressure [ mbar ] : 964  
 duct cross-section [ m<sup>2</sup> ] : 0.385

extraction parameter

change of meas. point [h:m:s] : 00:15:00  
 points / axis : 2  
 nozzles diameter [ mm ] : 10  
 isokinetic factor : 1  
 tare weight [ mg ] : 18457.17  
 gross weight [ mg ] : 18457.33

evaluation

meas. time [h:m:s] : 01:00:00  
 dust weight [ mg ] : 0.16

extracted partial volume  
 actual conditions [ m<sup>3</sup> ] : 1.614  
 in norm wet [ Nm<sup>3</sup> ] : 1.423  
 in norm dry [ Nm<sup>3</sup> ] : 1.390

volume flow in duct  
 actual conditions [m<sup>3</sup>/h ] : 7935  
 in norm wet [Nm<sup>3</sup>/h] : 6998  
 in norm dry [Nm<sup>3</sup>/h] : 6835

dust concentration  
 actual conditions [mg/m<sup>3</sup> ] : 0.10  
 in norm wet [mg/Nm<sup>3</sup> ] : 0.11  
 in norm dry [mg/Nm<sup>3</sup> ] : 0.12

protocol simultaneous isokinetic extraction measurement

11/11/2025 10:52

11/11/2025 11:54

measured values table

axis	depth	T_probe [°C]	v_duct [m/s]	angle [grd]	Q_act. [m³/h]	volume [m³]	meas. time [ H:M:S ]	p10 [mbar]	p40 [mbar]
1	1	22	5.8	1.0	1.62	0.406	00:15:00	1.00	-312
1	2	22	5.7	1.1	1.62	0.406	00:15:00	1.00	-311
2	1	22	5.7	0.8	1.60	0.402	00:15:00	1.00	-306
2	2	22	5.7	0.7	1.60	0.400	00:15:00	1.00	-307
		22	5.7	0.9	1.61	0.403		1.00	-309

protocol simultaneous isokinetic extraction measurement

11/11/2025 11:55

11/11/2025 12:57

collector-no. 1242  
 engineer PB DF  
 plant name Spray Boot LS  
 place CBE Plus  
 remarks Normal Operation  
 Dia = 700mm  
 T3

operating parameter

normal density humid [ kg / m<sup>3</sup> ] : 1.27  
 water vapour [ %Vol ] : 2.32  
 ambient pressure [ mbar ] : 965  
 duct cross-section [ m<sup>2</sup> ] : 0.385

extraction parameter

change of meas. point [h:m:s] : 00:15:00  
 points / axis : 2  
 nozzles diameter [ mm ] : 10  
 isokinetic factor : 1  
 tare weight [ mg ] : 18291.1  
 gross weight [ mg ] : 18291.24

evaluation

meas. time [h:m:s] : 01:00:00  
 dust weight [ mg ] : 0.14

extracted partial volume  
 actual conditions [ m<sup>3</sup> ] : 1.456  
 in norm wet [ Nm<sup>3</sup> ] : 1.284  
 in norm dry [ Nm<sup>3</sup> ] : 1.254

volume flow in duct  
 actual conditions [m<sup>3</sup>/h ] : 7138  
 in norm wet [Nm<sup>3</sup>/h] : 6295  
 in norm dry [Nm<sup>3</sup>/h] : 6149

dust concentration  
 actual conditions [mg/m<sup>3</sup> ] : 0.10  
 in norm wet [mg/Nm<sup>3</sup> ] : 0.11  
 in norm dry [mg/Nm<sup>3</sup> ] : 0.11

protocol simultaneous isokinetic extraction measurement

11/11/2025 11:55

11/11/2025 12:57

measured values table

axis	depth	T_probe [°C]	v_duct [m/s]	angle [grd]	Q_act. [m³/h]	volume [m³]	meas. time [ H:M:S ]	p10 [mbar]	p40 [mbar]
1	1	22	5.1	-5.1	1.45	0.362	00:15:00	1.00	-264
1	2	22	5.2	-5.3	1.46	0.366	00:15:00	1.00	-268
2	1	22	5.2	-5.6	1.45	0.364	00:15:00	1.00	-266
2	2	22	5.1	-5.5	1.46	0.364	00:15:00	1.00	-268
		22	5.1	-5.4	1.46	0.364		1.00	-267

**Site :** CBE Plus Limited  
**Date :** 11 November 2025  
**Plant :** Spray Booth Right Hand Side  
**File Ref.** 10720

**Mean Particulate Results**

Filter	Time	mg/m <sup>3</sup> (Actual Conditions)	m <sup>3</sup> /hr	mg/Nm <sup>3</sup> (Reference Conditions)	Nm <sup>3</sup> /hr	kg/hr
802010	09:50-10:52	0.25	10811	0.29	9534.0	0.003
80341	10:54-11:56	0.24	10776	0.27	9503.0	0.003
802039	12:00-13:02	0.50	10845	0.57	9479.0	0.005
<b>Mean</b>		0.33	10810.7	0.38	9505.3	0.004

**Control Blank Filter**

Filter	Volume (m <sup>3</sup> )	
802010	1.249	
80341	1.238	
802039	1.238	
<b>Mean</b>	1.242	(Reference Conditions with no correction for Oxygen)

**Filter** 115336

**Tare Weight** 18053.94 mg  
**Gross Weight** 18053.99 mg \*

**Gain** 0.04 mg

**Measured Oxygen** %

**Concentration** 0.04 mg/Nm<sup>3</sup>\*

Results Correct to

Temperature		Pressure		Oxygen		Gas	
°C/K	0/273	mbar/kPa	1013/101.3	%		Wet/Dry	Wet

\* Indicates where a value less than the limit of detection of the weighing procedure (0.09mg) has been reported, the value lies somewhere between the detection limit and zero. A value of half the limit of detection (0.045mg) has been used to calculate the concentration.

**Velocity Results**

	Time	m/s
<b>Test 1</b>	09:50-10:52	7.80
<b>Test 2</b>	10:54-11:56	7.80
<b>Test 3</b>	12:00-13:02	7.80
	<b>Mean</b>	<b>7.80</b>

protocol simultaneous isokinetic extraction measurement

11/11/2025 09:50

11/11/2025 10:52

collector-no. 2010  
 engineer PB DF  
 plant name Spray Boot RS  
 place CBE Plus  
 remarks Normal Operation  
 Dia = 700mm  
 T1

operating parameter

normal density humid [ kg / m<sup>3</sup> ] : 1.27  
 water vapour [ %Vol ] : 1.58  
 ambient pressure [ mbar ] : 964  
 duct cross-section [ m<sup>2</sup> ] : 0.385

extraction parameter

change of meas. point [h:m:s] : 00:15:00  
 points / axis : 2  
 nozzles diameter [ mm ] : 8  
 isokinetic factor : 1  
 tare weight [ mg ] : 18509.71  
 gross weight [ mg ] : 18510.07

evaluation

meas. time [h:m:s] : 01:00:00  
 dust weight [ mg ] : 0.36

extracted partial volume  
 actual conditions [ m<sup>3</sup> ] : 1.416  
 in norm wet [ Nm<sup>3</sup> ] : 1.249  
 in norm dry [ Nm<sup>3</sup> ] : 1.229

volume flow in duct  
 actual conditions [m<sup>3</sup>/h] : 10811  
 in norm wet [Nm<sup>3</sup>/h] : 9534  
 in norm dry [Nm<sup>3</sup>/h] : 9383

dust concentration  
 actual conditions [mg/m<sup>3</sup>] : 0.25  
 in norm wet [mg/Nm<sup>3</sup>] : 0.29  
 in norm dry [mg/Nm<sup>3</sup>] : 0.29

protocol simultaneous isokinetic extraction measurement

11/11/2025 09:50

11/11/2025 10:52

measured values table

axis	depth	T_probe [°C]	v_duct [m/s]	angle [grd]	Q_act. [m³/h]	volume [m³]	meas. time [ H:M:S ]	p10 [mbar]	p40 [mbar]
1	1	23	7.8	-1.4	1.43	0.358	00:15:00	1.00	-251
1	2	23	7.8	-0.9	1.40	0.350	00:15:00	1.00	-242
2	1	23	7.8	-1.2	1.41	0.354	00:15:00	1.00	-245
2	2	23	7.8	-0.8	1.42	0.354	00:15:00	1.00	-246
		23	7.8	-1.1	1.41	0.354		1.00	-246

protocol simultaneous isokinetic extraction measurement

11/11/2025 10:54

11/11/2025 11:56

collector-no. 341  
 engineer PB DF  
 plant name Spray Boot RS  
 place CBE Plus  
 remarks Normal Operation  
 Dia = 700mm  
 T2

operating parameter

normal density humid [ kg / m<sup>3</sup> ] : 1.27  
 water vapour [ %Vol ] : 1.58  
 ambient pressure [ mbar ] : 964  
 duct cross-section [ m<sup>2</sup> ] : 0.385

extraction parameter

change of meas. point [h:m:s] : 00:15:00  
 points / axis : 2  
 nozzles diameter [ mm ] : 8  
 isokinetic factor : 1  
 tare weight [ mg ] : 17605.35  
 gross weight [ mg ] : 17605.69

evaluation

meas. time [h:m:s] : 01:00:00  
 dust weight [ mg ] : 0.34

extracted partial volume  
 actual conditions [ m<sup>3</sup> ] : 1.404  
 in norm wet [ Nm<sup>3</sup> ] : 1.238  
 in norm dry [ Nm<sup>3</sup> ] : 1.219

volume flow in duct  
 actual conditions [m<sup>3</sup>/h ] : 10776  
 in norm wet [Nm<sup>3</sup>/h] : 9503  
 in norm dry [Nm<sup>3</sup>/h] : 9353

dust concentration  
 actual conditions [mg/m<sup>3</sup> ] : 0.24  
 in norm wet [mg/Nm<sup>3</sup> ] : 0.27  
 in norm dry [mg/Nm<sup>3</sup> ] : 0.28

protocol simultaneous isokinetic extraction measurement

11/11/2025 10:54

11/11/2025 11:56

measured values table

axis	depth	T_probe [°C]	v_duct [m/s]	angle [grd]	Q_act. [m³/h]	volume [m³]	meas. time [ H:M:S ]	p10 [mbar]	p40 [mbar]
1	1	23	7.8	2.5	1.40	0.352	00:15:00	1.00	-237
1	2	23	7.7	2.2	1.40	0.348	00:15:00	1.00	-236
2	1	23	7.8	2.2	1.40	0.350	00:15:00	1.00	-237
2	2	23	7.8	2.1	1.41	0.354	00:15:00	1.00	-241
		23	7.8	2.3	1.40	0.351		1.00	-238

protocol simultaneous isokinetic extraction measurement

11/11/2025 12:00

11/11/2025 13:02

collector-no. 2039  
 engineer PB DF  
 plant name Spray Boot RS  
 place CBE Plus  
 remarks Normal Operation  
 Dia = 700mm  
 T3

operating parameter

normal density humid [ kg / m<sup>3</sup> ] : 1.27  
 water vapour [ %Vol ] : 1.58  
 ambient pressure [ mbar ] : 964  
 duct cross-section [ m<sup>2</sup> ] : 0.385

extraction parameter

change of meas. point [h:m:s] : 00:15:00  
 points / axis : 2  
 nozzles diameter [ mm ] : 8  
 isokinetic factor : 1  
 tare weight [ mg ] : 18319.21  
 gross weight [ mg ] : 18319.92

evaluation

meas. time [h:m:s] : 01:00:00  
 dust weight [ mg ] : 0.71

extracted partial volume  
 actual conditions [ m<sup>3</sup> ] : 1.416  
 in norm wet [ Nm<sup>3</sup> ] : 1.238  
 in norm dry [ Nm<sup>3</sup> ] : 1.218

volume flow in duct  
 actual conditions [m<sup>3</sup>/h ] : 10845  
 in norm wet [Nm<sup>3</sup>/h] : 9479  
 in norm dry [Nm<sup>3</sup>/h] : 9329

dust concentration  
 actual conditions [mg/m<sup>3</sup> ] : 0.50  
 in norm wet [mg/Nm<sup>3</sup> ] : 0.57  
 in norm dry [mg/Nm<sup>3</sup> ] : 0.58

protocol simultaneous isokinetic extraction measurement

11/11/2025 12:00

11/11/2025 13:02

measured values table

axis	depth	T_probe [°C]	v_duct [m/s]	angle [grd]	Q_act. [m³/h]	volume [m³]	meas. time [ H:M:S ]	p10 [mbar]	p40 [mbar]
1	1	24	7.9	-0.4	1.43	0.356	00:15:00	1.00	-250
1	2	24	7.9	-0.4	1.43	0.358	00:15:00	1.00	-252
2	1	24	7.8	-0.4	1.41	0.352	00:15:00	1.00	-246
2	2	24	7.7	-0.6	1.40	0.350	00:15:00	1.00	-245
		24	7.8	-0.5	1.42	0.354		1.00	-248

**Appendix 4**  
**Gas (Instrument Measured Values & Results)**

Site	CBE Plus Ltd
Date	11/11/2025
Plant	Spray Booth LHS
File Ref	10720

Date & Time		VOC	
Start	End	wet ppm	wet mg/Nm <sup>3</sup>

11/11/2025 10:20	11/11/2025 11:19	71.36	114.69
------------------	------------------	-------	--------

Mean	71.36	114.69
Max	406.25	652.90
Min	11.41	18.33

Results Correct to

Temperature °C/K	Pressure mbar/kPa	Oxygen %	Gas Wet/Dry	Wet
0/273		1013/101.3	No Correction	

Date/Time	VOC Wet ppm	VOC Wet mg/Nm <sup>3</sup>
-----------	-------------------	----------------------------------

11/11/2025 10:20	15.94	25.61
11/11/2025 10:21	13.59	21.85
11/11/2025 10:22	13.44	21.60
11/11/2025 10:23	12.19	19.59
11/11/2025 10:24	12.19	19.59
11/11/2025 10:25	12.34	19.84
11/11/2025 10:26	12.19	19.59
11/11/2025 10:27	11.41	18.33
11/11/2025 10:28	76.72	123.30
11/11/2025 10:29	107.19	172.27
11/11/2025 10:30	116.09	186.58
11/11/2025 10:31	116.09	186.58
11/11/2025 10:32	116.09	186.58
11/11/2025 10:33	116.09	186.58
11/11/2025 10:34	116.09	186.58
11/11/2025 10:35	140.63	226.00
11/11/2025 10:36	123.44	198.38
11/11/2025 10:37	103.13	165.74
11/11/2025 10:38	84.38	135.60
11/11/2025 10:39	76.56	123.05
11/11/2025 10:40	70.31	113.00
11/11/2025 10:41	59.38	95.42
11/11/2025 10:42	54.69	87.89
11/11/2025 10:43	53.13	85.38
11/11/2025 10:44	46.88	75.33
11/11/2025 10:45	45.31	72.82
11/11/2025 10:46	40.63	65.29
11/11/2025 10:47	39.06	62.78
11/11/2025 10:48	37.50	60.27
11/11/2025 10:49	34.38	55.25
11/11/2025 10:50	32.81	52.73
11/11/2025 10:51	31.25	50.22
11/11/2025 10:52	29.69	47.71
11/11/2025 10:53	28.13	45.20
11/11/2025 10:54	29.69	47.71
11/11/2025 10:55	29.69	47.71
11/11/2025 10:56	26.56	42.69
11/11/2025 10:57	25.00	40.18
11/11/2025 10:58	75.00	120.54
11/11/2025 10:59	93.75	150.67
11/11/2025 11:00	120.31	193.36
11/11/2025 11:01	382.81	615.23
11/11/2025 11:02	406.25	652.90
11/11/2025 11:03	187.50	301.34
11/11/2025 11:04	120.31	193.36
11/11/2025 11:05	98.44	158.20
11/11/2025 11:06	87.50	140.63
11/11/2025 11:07	75.00	120.54
11/11/2025 11:08	64.06	102.96
11/11/2025 11:09	57.81	92.91
11/11/2025 11:10	48.44	77.85
11/11/2025 11:11	50.00	80.36
11/11/2025 11:12	46.88	75.33
11/11/2025 11:13	42.19	67.80
11/11/2025 11:14	40.63	65.29
11/11/2025 11:15	37.50	60.27
11/11/2025 11:16	39.06	62.78
11/11/2025 11:17	34.38	55.25
11/11/2025 11:18	31.25	50.22
11/11/2025 11:19	32.81	52.73
Min	11.41	18.33
Max	406.25	652.90
Mean	71.36	114.69

Site	CBE Plus Ltd
Date	11/11/2025
Plant	Spray Booth RHS
File Ref	10720

Date & Time		VOC	
Start	End	wet ppm	wet mg/Nm <sup>3</sup>

11/11/2025 11:35	11/11/2025 12:34	40.03	64.33
------------------	------------------	-------	-------

Mean	40.03	64.33
Max	187.50	301.34
Min	6.25	10.04

Results Correct to

Temperature °C/K	Pressure mbar/kPa	Oxygen %	Gas Wet/Dry	Wet
0/273	1013/101.3	No Correction		

Date/Time	VOC Wet ppm	VOC Wet mg/Nm <sup>3</sup>
-----------	-------------------	----------------------------------

11/11/2025 11:35	98.44	158.20
11/11/2025 11:36	142.19	228.52
11/11/2025 11:37	176.56	283.76
11/11/2025 11:38	96.88	155.69
11/11/2025 11:39	48.44	77.85
11/11/2025 11:40	39.06	62.78
11/11/2025 11:41	34.38	55.25
11/11/2025 11:42	25.00	40.18
11/11/2025 11:43	25.00	40.18
11/11/2025 11:44	20.31	32.65
11/11/2025 11:45	17.19	27.62
11/11/2025 11:46	18.75	30.13
11/11/2025 11:47	15.63	25.11
11/11/2025 11:48	18.75	30.13
11/11/2025 11:49	15.63	25.11
11/11/2025 11:50	14.06	22.60
11/11/2025 11:51	17.19	27.62
11/11/2025 11:52	14.06	22.60
11/11/2025 11:53	10.94	17.58
11/11/2025 11:54	10.94	17.58
11/11/2025 11:55	12.50	20.09
11/11/2025 11:56	10.94	17.58
11/11/2025 11:57	10.94	17.58
11/11/2025 11:58	9.38	15.07
11/11/2025 11:59	9.38	15.07
11/11/2025 12:00	9.38	15.07
11/11/2025 12:01	9.38	15.07
11/11/2025 12:02	7.81	12.56
11/11/2025 12:03	6.25	10.04
11/11/2025 12:04	9.38	15.07
11/11/2025 12:05	7.81	12.56
11/11/2025 12:06	9.38	15.07
11/11/2025 12:07	7.81	12.56
11/11/2025 12:08	6.25	10.04
11/11/2025 12:09	6.25	10.04
11/11/2025 12:10	7.81	12.56
11/11/2025 12:11	7.81	12.56
11/11/2025 12:12	6.25	10.04
11/11/2025 12:13	6.25	10.04
11/11/2025 12:14	7.81	12.56
11/11/2025 12:15	6.25	10.04
11/11/2025 12:16	17.19	27.62
11/11/2025 12:17	12.50	20.09
11/11/2025 12:18	76.56	123.05
11/11/2025 12:19	120.31	193.36
11/11/2025 12:20	139.06	223.49
11/11/2025 12:21	126.56	203.40
11/11/2025 12:22	160.94	258.65
11/11/2025 12:23	187.50	301.34
11/11/2025 12:24	154.69	248.60
11/11/2025 12:25	81.25	130.58
11/11/2025 12:26	54.69	87.89
11/11/2025 12:27	53.13	85.38
11/11/2025 12:28	39.06	62.78
11/11/2025 12:29	29.69	47.71
11/11/2025 12:30	29.69	47.71
11/11/2025 12:31	23.44	37.67
11/11/2025 12:32	21.88	35.16
11/11/2025 12:33	21.88	35.16
11/11/2025 12:34	17.19	27.62
Min	6.25	10.04
Max	187.50	301.34
Mean	40.03	64.33

**Appendix 5**  
**Instrumental Gas Analyser Site Calibration Data**

**CES Environmental Instruments Ltd**  
**GAS ANALYSER CALIBRATION SHEET**

Gas Analyser Calibration		Client	CBE Plus Ltd	Date	11/11/2025
Instrument Type	FID	Job Number	10720	Test	Auth
Quality No.	C301	Site	Spray Booth	Test Period	1 hour each stack

Range - Nitrogen (N2)	%
Range - Carbon Dioxide (CO2)	%
Range - Oxygen (O2)	%
Range - Carbon Monoxide (CO)	ppm
Range - Nitric Oxide (NO)	ppm
Range - Sulphur Dioxide (SO2)	ppm
Range - C <sub>3</sub> H <sub>8</sub>	1000.00 ppm

Zero Gas - Nitrogen (N <sub>2</sub> )	79 %	Certificate Number:	40-79302-001
Span Gas 1 - Carbon Dioxide (CO <sub>2</sub> )	%	Certificate Number:	
Span Gas 1 - Oxygen (O <sub>2</sub> )	%	Certificate Number:	
Span Gas 1 - Carbon Monoxide (CO)	ppm	Certificate Number:	
Span Gas 1 - Nitric Oxide (NO)	ppm	Certificate Number:	
Span Gas 1 - Sulphur Dioxide (SO <sub>2</sub> )	ppm	Certificate Number:	
Span Gas 1 - C <sub>3</sub> H <sub>8</sub>	808.9 ppm	Certificate Number:	40-96767-001

Span Gas 2 - Carbon Dioxide (CO <sub>2</sub> )	%	Certificate Number:	
Span Gas 2 - Oxygen (O <sub>2</sub> )	%	Certificate Number:	
Span Gas 2 - Carbon Monoxide (CO)	ppm	Certificate Number:	
Span Gas 2 - Nitric Oxide (NO)	ppm	Certificate Number:	
Span Gas 2 - Sulphur Dioxide (SO <sub>2</sub> )	ppm	Certificate Number:	
Span Gas 2 - C <sub>3</sub> H <sub>8</sub>	ppm	Certificate Number:	

Date	11/11/2025	Time		Amb Press (mbar)	990	Amb Temp. (°C)	10
------	------------	------	--	------------------	-----	----------------	----

Calibration Check - Pre-Sampling							
Test Gas Entered - Instrument	CO <sub>2</sub>	O <sub>2</sub>	CO	NO	SO <sub>2</sub>	C3H8	
Zero Reading							0.00
Span Gas 1 Reading							809.00
Zero Reading							1.00
Ambient Air							
Analyser T <sub>90</sub> Response (sec)							11.00
<2x Repeatability @ zero	0.00	0.00	0.00	0.00	0.00	0.00	1.00

Leak Check - Pre-Sampling							
Test Gas Entered - Probe	CO <sub>2</sub>	O <sub>2</sub>	CO	NO	SO <sub>2</sub>	C3H8	
Zero Reading							1.00
Span Gas 1 Reading							810.00
Span Gas 2 Reading							
Ambient Air							
Zero Drift (%)							0.10
Span Drift (%)							0.10

Date	11/11/2025	Time		Amb Press (mbar)	990	Amb Temp. (°C)	11
------	------------	------	--	------------------	-----	----------------	----

Calibration Check - Mid / Post Sampling							
Test Gas Entered - Instrument	CO <sub>2</sub>	O <sub>2</sub>	CO	NO	SO <sub>2</sub>	C3H8	
Zero Reading							1.00
Span Gas 1 Reading							811.00
Span Gas 2 Reading							
Ambient Air							
Zero Drift (%)							0.10
Span Drift (%)							0.20

Checks completed by: P.Broxham Page 1 of 1

**CES Environmental Instruments Ltd**  
**GAS ANALYSER CALIBRATION SHEET**

Gas Analyser Calibration		Client	CBE Plus Ltd	Date	11/11/2025
Instrument Type	FID	Job Number	10720	Test	Auth
Quality No.	C301	Site	Spray Booth	Test Period	1 hour each stack

Range - Nitrogen (N2)	%
Range - Carbon Dioxide (CO2)	%
Range - Oxygen (O2)	%
Range - Carbon Monoxide (CO)	ppm
Range - Nitric Oxide (NO)	ppm
Range - Sulphur Dioxide (SO2)	ppm
Range - C <sub>3</sub> H <sub>8</sub>	1000.00 ppm

Zero Gas - Nitrogen (N <sub>2</sub> )	79 %	Certificate Number:	40-79302-001
Span Gas 1 - Carbon Dioxide (CO <sub>2</sub> )	%	Certificate Number:	
Span Gas 1 - Oxygen (O <sub>2</sub> )	%	Certificate Number:	
Span Gas 1 - Carbon Monoxide (CO)	ppm	Certificate Number:	
Span Gas 1 - Nitric Oxide (NO)	ppm	Certificate Number:	
Span Gas 1 - Sulphur Dioxide (SO <sub>2</sub> )	ppm	Certificate Number:	
Span Gas 1 - C <sub>3</sub> H <sub>8</sub>	808.9 ppm	Certificate Number:	40-96767-001

Span Gas 2 - Carbon Dioxide (CO <sub>2</sub> )	%	Certificate Number:	
Span Gas 2 - Oxygen (O <sub>2</sub> )	%	Certificate Number:	
Span Gas 2 - Carbon Monoxide (CO)	ppm	Certificate Number:	
Span Gas 2 - Nitric Oxide (NO)	ppm	Certificate Number:	
Span Gas 2 - Sulphur Dioxide (SO <sub>2</sub> )	ppm	Certificate Number:	
Span Gas 2 - C <sub>3</sub> H <sub>8</sub>	ppm	Certificate Number:	

Date	11/11/2025	Time		Amb Press (mbar)	990	Amb Temp. (°C)	10
------	------------	------	--	------------------	-----	----------------	----

Calibration Check - Pre-Sampling							
Test Gas Entered - Instrument	CO <sub>2</sub>	O <sub>2</sub>	CO	NO	SO <sub>2</sub>	C3H8	
Zero Reading							0.00
Span Gas 1 Reading							809.00
Zero Reading							1.00
Ambient Air							
Analyser T <sub>90</sub> Response (sec)							11.00
<2x Repeatability @ zero	0.00	0.00	0.00	0.00	0.00	0.00	1.00

Leak Check - Pre-Sampling							
Test Gas Entered - Probe	CO <sub>2</sub>	O <sub>2</sub>	CO	NO	SO <sub>2</sub>	C3H8	
Zero Reading							1.00
Span Gas 1 Reading							810.00
Span Gas 2 Reading							
Ambient Air							
Zero Drift (%)							0.10
Span Drift (%)							0.10

Date	11/11/2025	Time		Amb Press (mbar)	990	Amb Temp. (°C)	11
------	------------	------	--	------------------	-----	----------------	----

Calibration Check - Mid / Post Sampling							
Test Gas Entered - Instrument	CO <sub>2</sub>	O <sub>2</sub>	CO	NO	SO <sub>2</sub>	C3H8	
Zero Reading							1.00
Span Gas 1 Reading							811.00
Span Gas 2 Reading							
Ambient Air							
Zero Drift (%)							0.10
Span Drift (%)							0.20

Checks completed by: P.Broxham Page 1 of 1

**Appendix 6**  
**Calibration Certificates**

## Certificate of Calibration

Date of Issue: 13th June 2025

Certificate No. CES2638  
page 1 of 2

CES Environmental Instruments Ltd  
Bretby Business Park, Ashby Road  
Burton-on-Trent, Staffordshire, DE15 0YZ  
Tel: 01283 216334 Fax: 01283 550939

  
\_\_\_\_\_  
Certified By

### Instrument Details

Instrument Type	Gravimat SHC-502
Instrument Make	Erwin Sick
Instrument Serial No.	14288454
Quality No.	C300
Calibration Date	13 06 25
Calibrated By Name	M.Rodgers

### Ambient Conditions

Air Temperature (°C)	25
Barometric Pressure (mbar)	1003
Relative Humidity (%)	65

### Instruments used to undertake calibration

E Type Pitot	UKAS Certificate No. K55647V	(Qu. No. C136)
Manometer Type FC012	UKAS Certificate No. 27318	(Qu. No. C082)
Manometer Type FC012	UKAS Certificate No. 27319	(Qu. No. C081)
Barometer Type 104	UKAS Certificate No. K55933P	(Qu. No. C138)
Gallus Dry Gas Meter	UKAS Certificate No. N039726	(Qu. No. C333)
RIS Supercal XT	UKAS Certificate No. 10345	(Qu. No. C014)

### Flow and Extraction

The reference pitot was placed in a wind tunnel located at Bretby Business Park. The Gravimat SHC-502 Sampling Probe under test was mounted within the same wind tunnel in close proximity to the reference pitot. The wind tunnel was operated to generate a differential pressure across each pitot, a direct comparison was made. The differential pressures measured were in the region of the calibration points of the reference pitot. Correction factors were applied to the reference pitot and compared to the differential pressure shown for the pitot under test. The extraction system of the unit was operated for a period of one minute.

### Volume Flow

A calibrated dry gas meter was connected to the sampling nozzle of the Gravimat SCH-502. A volume of air is pulled through the sampling system. The measured value shown on the calibrated dry gas meter is then compared to the indicated value on the Gravimat SCH-5 display.

### Barometric Pressure

The barometric pressure was measured using a calibrated barometer. The indicated pressure was compared to the Gravimat SHC-502 display.

### Temperature

The probe thermocouple was placed in a thermocouple oven and heated. The temperature was measured using a calibrated thermocouple and temperature indicator. The resultant temperature was compared to the Gravimat SCH-502 display.

### Current

A mA current source was injected into the Gravimat SCH-502 using a mA current generator. The injected current was compared to the Gravimat SCH-502 display.

## Certificate of Calibration

Date of Issue: 13th June 2025

Certificate No CES2638  
page 2 of 2

CES Environmental Instruments Ltd  
Bretby Business Park, Ashby Road  
Burton-on-Trent, Staffordshire, DE15 0YZ  
Tel: 01283 216334 Fax: 01283 550939

  
Certified By

### Instrument Details

Instrument Type Gravimat SHC-502  
Instrument Make Erwin Sick  
Instrument Serial No 14288454  
Quality No C300  
Calibration Date 13 06 25

### Ambient Conditions

Air Temperature (°C) 25  
Barometric Pressure (mbar) 1003  
Relative Humidity (%) 65  
Air Density @ 0°C (kg/m³) 1.277  
Corrected Air Density (kg/m³) 1.1583

### Calibration Details

#### Flow and Extraction

	Pitot Correction	Applied Pressure Corrected (Pa)	SHC502 (Calculated) (Pa)	Pressure Factor	Calculated Velocity (m/s)	SHC502 Velocity (m/s)	Velocity Factor
7.1	1.009	7.2	7.5	0.96	3.526	3.600	0.98
45.0	1.002	45.1	46.7	0.97	8.846	9.000	0.98
89.5	1.004	89.9	91.5	0.98	12.487	12.600	0.99
149.6	0.999	149.5	156.9	0.95	16.104	16.500	0.98
235.9	1	235.9	249.3	0.95	20.233	20.800	0.97
<b>Mean (excluding 4.0Pa)</b>				<b>0.96</b>			<b>0.98</b>

#### Volume Flow

Nominal Flow Rate l/min	Actual Flow Rate l/min	Actual Flow Rate m³/hr	Orifice Constant
10.0	9.560	0.191	12.894
15.0	14.740	0.295	13.250
20.0	20.495	0.410	13.440
25.0	22.570	1.354	13.483
30.0	30.055	0.601	13.462
<b>Mean Orifice Constant(@1.5m³/hr) =</b>			<b>13.483</b>

#### Barometric Pressure

Pressure Input	
Required Value (mbar)	Indicated Value (mbar)
995.0	995.0
1002.0	1002.0
1007.0	1007.0

#### Temperature

Temperature Input	
Required Value (°C)	Indicated Value (°C)
25.0	25.0
50.0	50.0
100.0	100.0
150.0	150.0
250.0	251.0
300.0	301.0

#### Current

Current Value	
Required Value (mA)	Indicated Value (mA)
0	0.0
5	5.0
10	10.0
15	15.0
20	20.0

#### Time

Time Period mins	Required Value mins	Within Limit
3:00	2:59 → 3:01	Yes
5:00	4:59 → 5:01	Yes
10:00	9:59 → 10:01	Yes

## Certificate of Calibration

Date of Issue: 6th January 2025

Certificate No. CES2583  
page 1 of 2

CES Environmental Instruments Ltd  
Bretby Business Park, Ashby Road  
Burton-on-Trent, Staffordshire, DE15 0YZ  
Tel: 01283 216334 Fax: 01283 550939

  
Certified By

### Instrument Details

Instrument Type	Gravimat SHC-502
Instrument Make	Erwin Sick
Instrument Serial No.	5218792
Quality No.	C193
Calibration Date	06 01 2025
Calibrated By Name	D.Woolley

### Ambient Conditions

Air Temperature (°C)	23
Barometric Pressure (mbar)	987
Relative Humidity (%)	44

### Instruments used to undertake calibration

E Type Pitot	UKAS Certificate No. K55647V	(Qu. No. C136)
Manometer Type FC012	UKAS Certificate No. 27318	(Qu. No. C082)
Manometer Type FC012	UKAS Certificate No. 27319	(Qu. No. C081)
Barometer Type 104	UKAS Certificate No. K55933P	(Qu. No. C138)
Gallus Dry Gas Meter	UKAS Certificate No. N037163	(Qu. No. C333)
RIS Supercal XT	UKAS Certificate No. 10345	(Qu. No. C014)

### Flow and Extraction

The reference pitot was placed in a wind tunnel located at Bretby Business Park. The Gravimat SHC-5 Sampling Probe under test was mounted within the same wind tunnel in close proximity to the reference pitot. The wind tunnel was operated to generate a differential pressure across each pitot, a direct comparison was made. The differential pressures measured were in the region of the calibration points of the reference pitot. Correction factors were applied to the reference pitot and compared to the differential pressure shown for the pitot under test. The extraction system of the unit was operated for a period of one minute.

### Volume Flow

A calibrated dry gas meter was connected to the sampling nozzle of the Gravimat SCH-5. A volume of air is pulled through the sampling system. The measured value shown on the calibrated dry gas meter is then compared to the indicated value on the Gravimat SCH-5 display.

### Barometric Pressure

The barometric pressure was measured using a calibrated barometer. The indicated pressure was compared to the Gravimat SHC-5 display.

### Temperature

The probe thermocouple was placed in a thermocouple oven and heated. The temperature was measured using a calibrated thermocouple and temperature indicator. The resultant temperature was compared to the Gravimat SCH-5 display.

### Current

A mA current source was injected into the Gravimat SCH-5 using a mA current generator. The injected current was compared to the Gravimat SCH-5 display.

## Certificate of Calibration

Date of Issue: 6th January 2025

Certificate No. CES2583  
page 2 of 2

CES Environmental Instruments Ltd  
Bretby Business Park, Ashby Road  
Burton-on-Trent, Staffordshire, DE15 0YZ  
Tel: 01283 216334 Fax: 01283 550939

  
Certified By

### Instrument Details

Instrument Type Gravimat SHC-502  
Instrument Make Erwin Sick  
Instrument Serial No. 5218792  
Quality No C193  
Calibration Date 06 01 2025

### Ambient Conditions

Air Temperature (°C) 23  
Barometric Pressure (mbar) 987  
Relative Humidity (%) 44  
Air Density @ 0°C (kg/m³) 1.277  
Corrected Air Density (kg/m³) 1.203

### Calibration Details

#### Flow and Extraction

Applied Pressure (Pa)	Pitot Correction	Applied Pressure Corrected (Pa)	SMP502 (Calculated) (Pa)	Pressure Factor	Calculated Velocity (m/s)	SMP502 Velocity (m/s)	Velocity Factor
11.0	1.018	11.20	10.5	1.07	4.314	4.177	1.03
37.0	1.005	37.19	35.5	1.05	7.861	7.681	1.02
84.0	1.007	84.59	81.5	1.04	11.857	11.639	1.02
149.0	0.998	148.70	144.5	1.03	15.721	15.497	1.01
235.0	0.999	234.77	226.5	1.04	19.753	19.402	1.02
363.0	0.982	356.47	349.5	1.02	24.341	24.102	1.01
523.0	0.983	514.11	504.5	1.02	29.231	28.957	1.01

#### Volume Flow

Nominal Flow Rate	Actual Flow Rate	Actual Flow Rate	Orifice Constant
l/min	l/min	m³/hr	
10.0	9.475	0.190	12.536
15.0	14.720	0.294	13.200
20.0	19.830	0.397	13.335
25.0	23.890	1.433	13.449
30.0	30.600	0.612	13.719
ment Orifice Constant (@1.5m³/hr) =			13.449

### Barometric Pressure

Pressure	
Required Value (mbar)	Indicated Value (mbar)
987.0	987.0
1000.0	1000.0
1021.0	1021.0

### Temperature

Temperature Input	
Required Value (°C)	Indicated Value (°C)
25.0	25.0
50.0	49.9
100.0	100
150.0	150.2
250.0	250.1
300.0	300.0

### Current

Current Value	
Required Value (mA)	Indicated Value (mA)
0.0	0.0
5.0	5.0
10.0	10.0
15.0	15.0
20.0	20.0

### Time

Time Period mins	Required Value mins	Within Limit
3:00	2:59 → 3:01	Yes
5:00	4:59 → 5:01	Yes
10:00	9:59 → 10:01	Yes

## **Appendix 7**

### **Uncertainty Calculations**

Uncertainty calculation for EN 13284 Determination of low range mass concentration of dust, Manual Gravimetric Method

Measurement Equation

$$c = \frac{m}{V} f_c$$

Limit value (ELV)		mg.m <sup>-3</sup>	Reference oxygen		% by volume
Measured concentration	0.11	mg.m <sup>-3</sup> (at reference conditions)			

Measured Quantities	Symbol	Value	Standard uncertainty	Units	Uncertainty as percentage	Uncertainty at lv	Requirement of std
Sampled Volume	V <sub>m</sub>	1.423	uV <sub>m</sub>	0.001 m <sup>3</sup>	0.07		<=2%
Sampled gas Temperature	T <sub>m</sub>	273	uT <sub>m</sub>	2 k	0.73		<=1%
Sampled gas Pressure	ρ <sub>m</sub>	101.3	uρ <sub>m</sub>	0.1 kPa	0.10		<=1%
Sampled gas Humidity	H <sub>m</sub>	0	uH <sub>m</sub>	1 % by volume	1.00		<=1%
Oxygen content	O <sub>2,m</sub>		uO <sub>2,m</sub>	0.1 % by volume	#DIV/0!		<=5%
Mass particulate	m	0.16	um	0.09 mg	56.25	#DIV/0!	<5% of limit value
Note - Sampled gas humidity, temperature and pressure are values at the gas meter							
Leak	L	2		%	2.00		<=2%
Uncollected Mass (Instack filter - no rinse)	UCM	0.01		mg	6.25		<=10%

Intermediate calculations				
Factor for std conds	fs	1.00		
uncertainty components	symbol	sensitivity coeff	u (in units of fs)	
	ρ <sub>m</sub>	0.010	0.001	
	H <sub>m</sub>	0.010	0.010	$f_s = \frac{(100 - H_m) 273 \rho_m}{100 T_m 101.3}$
	T <sub>m</sub>	0.004	0.007	
	ufs		0.012	1.24
Corrected volume	V	1.42	uV	0.018 m <sup>3</sup>
				1.25
Factor for O2 correction	fc	1.00		
uncertainty components	symbol	sensitivity coeff	u	
	O <sub>2,m</sub>	0.05	0.005	$f_c = \frac{21 - O_{2,ref}}{21 - O_{2,m}}$
Factor for O2 Correction	ufc	1.00	0.005	0.48

Parameter	Value	Units	Sensitivity coeff	Uncertainty contribution	Uncertainty as %
Corrected Volume (standard cc)	V	1.42 m <sup>3</sup>	0.08	0.00 mg.m <sup>-3</sup>	1.25 %
Mass	m	0.16 mg	0.70	0.06 mg.m <sup>-3</sup>	56.25 %
Factor for O2 Correction	fc	1.00	0.11	0.00 mg.m <sup>-3</sup>	0.48 %
Leak	L	0.00 mg.m <sup>-3</sup>	1.00	0.00 mg.m <sup>-3</sup>	1.15 %
Uncollected mass	UCM	0.01 mg	0.70	0.00 mg.m <sup>-3</sup>	3.61 %
<b>Combined measurement uncertainty</b>				<b>0.06 mg.m<sup>-3</sup></b>	
Expanded uncertainty as percentage of measured value		112.79	% measured of value	expressed with a level of confidence of 95% (Using a coverage factor k=2)	
Expanded uncertainty in units of measurement		0.13	mg.m <sup>-3</sup>		
Expanded uncertainty as percentage of limit value		#DIV/0!	% ELV		

Verified

Standard uncertainty associated with local velocities (BS EN ISO 16911-1:2013)

Measured velocity	5.80 m/s	O2 concentration	20.9 % dry
Measured temperature	22.00 °C	CO2 Concentration	0.03 %dry
Measured ambient pressure	96300.00 Pa	H2O Concentration	2.32 %
Duct Pressure, Abs	96400.00 Pa	Mole Mass	28.589 kg/mol
Differential Duct Press	100.00 Pa	Gas Density	1.276 kg/m³(0°C, 1013mbar) 1.124 kg/m³(actual)
Measurement points	4		
<b>Allowable uncertainties</b>			
Max Allowable U Dry O <sub>2</sub> % v/v =	6	Accuracy of absolute temp % 1	
Max Allowable U Dry CO <sub>2</sub> % v/v =	6	Coverage @ 95% CI 2	
Max Allowable U H <sub>2</sub> O % v/v =	20		
<b>Measured</b>			
	Min	Max	Mean
Local static pressure (Pa)	100.00	100.00	100.00
Local dynamic pressure (Pa)	18.20	18.80	18.80
<b>Relating to ambient pressure (Pa)</b>			
	Est Reading Error	Max permissible error	
	20	300	
<b>Instrumentation</b>			
Temperature measurement: K	295.00	Accuracy +/-	2.95
Pitot tube calibration factor	1	Tolerance +/-	0.02
Pressure indicator range: Pa zero to	1000.00	Calibration U Pa +/-	2.00
		Resolution (Pa)	1.00
		Drift (%)	0.10
		Lack of Fit (% of Range)	0.06
<b>Gas composition</b>			
		MW	
Dry N <sub>2</sub> % v/v =	79.070	28.00	Wet N <sub>2</sub> % v/v = 79.556
Dry O <sub>2</sub> % v/v =	20.900	32.00	Wet O <sub>2</sub> % v/v = 20.415
Dry CO <sub>2</sub> % v/v =	0.030	44.00	Wet CO <sub>2</sub> % v/v = 0.029
Total %	100.000		Total % 100.000
H <sub>2</sub> O =	2.320	18.00	
<b>Calculated uncertainties</b>			
	Value		
Standard uncertainty associated with the mean local dynamic pressures	1.25	Pa	
Standard uncertainty on the coefficient of the pitot tube	0.01		
Uncertainty associated with H <sub>2</sub> O	0.23	% volume fraction on wet gas	
Uncertainty associated with O <sub>2</sub>	0.61	% volume fraction water vapour	
Uncertainty associated with CO <sub>2</sub>	0.00	% volume fraction water vapour	
Standard uncertainty associated with the molar mass of gas	3.38E-05	kg/mol	
Standard uncertainty associated with temperature	1.48	K	
Standard uncertainty associated with ambient pressure	173.30	Pa	
Standard uncertainty associated with static pressure	1.24	Pa	
Standard uncertainty associated with density	6.12E-03	kg/m³	
<b>Measurement uncertainty</b>			
<b>Calculation of uncertainty associated with the determination of local velocities</b>			
Expanded uncertainty	k = 2	0.20	m/s
		0.40	m/s

Uncertainty calculation for Gaseous Measurement Volatile Organic Compounds

BS EN 12619

Limit value	mg.m <sup>-3</sup> (corrected)	Gas	C3H8
Uncertainty Requirement	15 %	Full Scale	1000 ppm
Measured concentration	71.36 ppm	Cal gas conc	808.9 ppm
Measured concentration	114.18 mgC.m <sup>-3</sup> (273K, 101.3kPa)	Conversion	1.6
Measured concentration	117.71 mgC.m <sup>-3</sup> (corrected)	Full Scale	1600 mg.m <sup>-3</sup>
		Cal gas conc	1294.24 mg.m <sup>-3</sup>

Correction for reference conditions				
	O2, %	Moisture, %	Pressure, KPa	Temperature, K
ref	0.00	0.00	101.30	273.00
measured	0.00	3.00	101.30	273.00
Uncert	0.05			1.00
Factors	1.00	1.03	1.00	1.00
Uncertainty in factor	0.00	0.00		0.00
Correction Factor	1.03	uf	0.00436881334867	

Performance characteristics	Value		specification
Response time	11.00	seconds	180.000
Number of readings in measurement	60		
Repeatability at zero	0.15	% full scale	0.200
Repeatability at span level	0.8	% full scale	2.000
Deviation from linearity	0.4	% of value	2.000
Zero drift	0.10	% full scale	2.000
Span drift	0.20	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0.5	% of value/kPa	0.750
ambient temperature dependence	0.01	% full scale/10K	0.300
N2O (40 mg/m3)	0.3	mg/m3	
CH4 (57 mg/m3)	0.3	mg/m3	
NH3 (20 mg/m3)	0.3	mg/m3	
CO2 (15%)	0.3	% by vol	
H2O (30%)	0.5	% by vol	4.000
dependence on voltage	0.1	% full scale/10V	2%/s/10V
losses in the line (leak)	2	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift	
	1.83 mg/m3
	0.11 % value

	ranges		
	min	max	value at calib
flow	18	30	24 l/hr
pressure	99.00	99.00	99.00 kPa
temp	283.00	284.00	283.00 K
NH3 range	0	20	0 mg/m3
N2O range	0	40	0 mg/m3
CH4 range	0	57	0 mg/m3
CO2 range	8	10	0 %vol
H2O range	3.00	3.00	0 %vol
Instrument Voltage Rating	110 V		
Voltage	104.5	115.5	110 V

Measurement performance related to stationary conditions				
Performance characteristic	Uncertainty	Value of uncertainty quantity		
		for mean		use rep at span
Standard deviation of repeatability at zero	U <sub>r0</sub>			
Standard deviation of repeatability at span level	U <sub>rs</sub>			0.10
Lack of fit	U <sub>fit</sub>			3.70
Drift	U <sub>odr</sub>			1.06
volume or pressure flow dependence	U <sub>spres</sub>			0.11
atmospheric pressure dependence	U <sub>apres</sub>			0.00
ambient temperature dependence	U <sub>tamp</sub>			0.00
NH3 (20 mg/m3)	U <sub>interf</sub>			0.17
CO2 (15%)	U <sub>interf</sub>			0.18
H2O (30%)	U <sub>interf</sub>			0.05
N2O (40 mg/m3)	U <sub>interf</sub>			0.17
CH4 (57 mg/m3)	U <sub>interf</sub>			0.17
Dependence on voltage	U <sub>volt</sub>			0.03
losses in the line (leak)	U <sub>leak</sub>			1.32
Uncertainty of calibration gas	U <sub>calib</sub>			1.32
Uncertainty in factor	uf			0.00

Use largest negative or positive interferent effect		
NH3	0	0.17
CO2	0	0.18
H2O	0	0.05
N2O	0	0.17
CH4	0	0.17
	0	0.75
Interference uncertainty		0.75

Measurement uncertainty	Result		
Combined uncertainty		114.18	mg/m <sup>3</sup>
Expanded uncertainty	k = 2	4.34	mg/m <sup>3</sup>
Expanded uncertainty		8.69	mg/m <sup>3</sup>
Uncertainty corrected to std conds		8.95	mg.m-3 (corrected)
Expanded uncertainty	expressed with a level of confidence of 95%	#VALUE!	% ELV
Expanded uncertainty	expressed with a level of confidence of 95%	8.95	mg.m <sup>-3</sup> at ELV
	U <sub>95</sub> ≤ U <sub>req</sub>	#VALUE!	

Verified

Uncertainty calculation for EN 13284 Determination of low range mass concentration of dust, Manual Gravimetric Method

Measurement Equation

$$c = \frac{m}{V} f_c$$

Limit value (ELV)		mg.m <sup>-3</sup>	Reference oxygen		% by volume
Measured concentration	0.57	mg.m <sup>-3</sup> (at reference conditions)			

Measured Quantities	Symbol	Value	Standard uncertainty	Units	Uncertainty as percentage	Uncertainty at lv	Requirement of std
Sampled Volume	V <sub>m</sub>	1.238	uV <sub>m</sub>	0.001 m <sup>3</sup>	0.08		<=2%
Sampled gas Temperature	T <sub>m</sub>	273	uT <sub>m</sub>	2 k	0.73		<=1%
Sampled gas Pressure	ρ <sub>m</sub>	101.3	uρ <sub>m</sub>	0.1 kPa	0.10		<=1%
Sampled gas Humidity	H <sub>m</sub>	0	uH <sub>m</sub>	1 % by volume	1.00		<=1%
Oxygen content	O <sub>2,m</sub>		uO <sub>2,m</sub>	0.1 % by volume	#DIV/0!		<=5%
Mass particulate	m	0.71	um	0.09 mg	12.68	#DIV/0!	<5% of limit value
Note - Sampled gas humidity, temperature and pressure are values at the gas meter							
Leak	L	2		%	2.00		<=2%
Uncollected Mass (Instack filter - no rinse)	UCM	0.05		mg	7.042253521		<=10%

Intermediate calculations				
Factor for std conds	fs	1.00		
uncertainty components	symbol	sensitivity coeff	u (in units of fs)	
	ρ <sub>m</sub>	0.010	0.001	
	H <sub>m</sub>	0.010	0.010	$f_s = \frac{(100 - H_m) 273 \rho_m}{100 T_m 101.3}$
	T <sub>m</sub>	0.004	0.007	
	ufs		0.012	1.24
Corrected volume	V	1.24	uV	0.015 m <sup>3</sup>
				$V = V_m f_s$
Factor for O2 correction	fc	1.00		
uncertainty components	symbol	sensitivity coeff	u	
	O <sub>2,m</sub>	0.05	0.005	$f_c = \frac{21 - O_{2,ref}}{21 - O_{2,m}}$
Factor for O2 Correction	ufc	1.00	0.005	0.48

Parameter	Value	Units	Sensitivity coeff	Uncertainty contribution	Uncertainty as %
Corrected Volume (standard cc	V	1.24 m <sup>3</sup>	0.46	0.01 mg.m <sup>-3</sup>	1.25 %
Mass	m	0.71 mg	0.81	0.07 mg.m <sup>-3</sup>	12.68 %
Factor for O2 Correction	fc	1.00	0.57	0.00 mg.m <sup>-3</sup>	0.48 %
Leak	L	0.01 mg.m <sup>-3</sup>	1.00	0.01 mg.m <sup>-3</sup>	1.15 %
Uncollected mass	UCM	0.03 mg	0.81	0.02 mg.m <sup>-3</sup>	4.07 %
<b>Combined measurement uncertainty</b>				<b>0.08 mg.m<sup>-3</sup></b>	
Expanded uncertainty as percentage of measured value		26.86	% measured of value	expressed with a level of confidence of 95% (Using a coverage factor k=2)	
Expanded uncertainty in units of measurement		0.15	mg.m <sup>-3</sup>		
Expanded uncertainty as percentage of limit value		#DIV/0!	% ELV		

Verified

Standard uncertainty associated with local velocities (BS EN ISO 16911-1:2013)

Measured velocity	7.80 m/s	O2 concentration	20.9 % dry
Measured temperature	24.00 °C	CO2 Concentration	0.03 %dry
Measured ambient pressure	96100.00 Pa	H2O Concentration	1.58 %
Duct Pressure, Abs	96200.00 Pa	Mole Mass	28.670 kg/mol
Differential Duct Press	100.00 Pa	Gas Density	1.280 kg/m³(0°C, 1013mbar)
Measurement points	4		1.117 kg/m³(actual)
<b>Allowable uncertainties</b>			
Max Allowable U Dry O <sub>2</sub> % v/v =	6	Accuracy of absolute temp % 1	
Max Allowable U Dry CO <sub>2</sub> % v/v =	6	Coverage @ 95% CI 2	
Max Allowable U H <sub>2</sub> O % v/v =	20		
<b>Measured</b>			
	Min	Max	Mean
Local static pressure (Pa)	100.00	100.00	100.00
Local dynamic pressure (Pa)	32.80	34.60	33.70
<b>Relating to ambient pressure (Pa)</b>			
	Est Reading Error	Max permissible error	
	20	300	
<b>Instrumentation</b>			
Temperature measurement: K	297.00	Accuracy +/-	2.97
Pitot tube calibration factor	1	Tolerance +/-	0.02
Pressure indicator range: Pa zero to	1000.00	Calibration U Pa +/-	2.00
		Resolution (Pa)	1.00
		Drift (%)	0.10
		Lack of Fit (% of Range)	0.06
<b>Gas composition</b>			
		MW	
Dry N <sub>2</sub> % v/v =	79.070	28.00	Wet N <sub>2</sub> % v/v = 79.401
Dry O <sub>2</sub> % v/v =	20.900	32.00	Wet O <sub>2</sub> % v/v = 20.570
Dry CO <sub>2</sub> % v/v =	0.030	44.00	Wet CO <sub>2</sub> % v/v = 0.030
Total %	100.000		Total % 100.000
H <sub>2</sub> O =	1.580	18.00	
<b>Calculated uncertainties</b>			
	Value		
Standard uncertainty associated with the mean local dynamic pressures	1.31	Pa	
Standard uncertainty on the coefficient of the pitot tube	0.01		
Uncertainty associated with H <sub>2</sub> O	0.16	% volume fraction on wet gas	
Uncertainty associated with O <sub>2</sub>	0.62	% volume fraction water vapour	
Uncertainty associated with CO <sub>2</sub>	0.00	% volume fraction water vapour	
Standard uncertainty associated with the molar mass of gas	2.93E-05	kg/mol	
Standard uncertainty associated with temperature	1.49	K	
Standard uncertainty associated with ambient pressure	173.30	Pa	
Standard uncertainty associated with static pressure	1.24	Pa	
Standard uncertainty associated with density	6.05E-03	kg/m³	
<b>Measurement uncertainty</b>			
<b>Calculation of uncertainty associated with the determination of local velocities</b>		0.17	m/s
Expanded uncertainty	k = 2	0.34	m/s

Uncertainty calculation for Gaseous Measurement Volatile Organic Compounds

BS EN 12619

Limit value	mg.m <sup>-3</sup> (corrected)	Gas	C3H8
Uncertainty Requirement	15 %	Full Scale	1000 ppm
Measured concentration	40.03 ppm	Cal gas conc	808.9 ppm
Measured concentration	64.04 mgC.m <sup>-3</sup> (273K, 101.3kPa)	Conversion	1.6
Measured concentration	66.02 mgC.m <sup>-3</sup> (corrected)	Full Scale	1600 mg.m <sup>-3</sup>
		Cal gas conc	1294.24 mg.m <sup>-3</sup>

Correction for reference conditions				
	O2, %	Moisture, %	Pressure, KPa	Temperature, K
ref	0.00	0.00	101.30	273.00
measured	0.00	3.00	101.30	273.00
Uncert	0.05			1.00
Factors	1.00	1.03	1.00	1.00
Uncertainty in factor	0.00	0.00		0.00
Correction Factor	1.03	uf	0.00436881334867	

Performance characteristics	Value		specification
Response time	11.00	seconds	180.000
Number of readings in measurement	60		
Repeatability at zero	0.15	% full scale	0.200
Repeatability at span level	0.8	% full scale	2.000
Deviation from linearity	0.4	% of value	2.000
Zero drift	0.10	% full scale	2.000
Span drift	0.20	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0.5	% of value/kPa	0.750
ambient temperature dependence	0.01	% full scale/10K	0.300
N2O (40 mg/m3)	0.3	mg/m3	
CH4 (57 mg/m3)	0.3	mg/m3	
NH3 (20 mg/m3)	0.3	mg/m3	
CO2 (15%)	0.3	% by vol	
H2O (30%)	0.5	% by vol	4.000
dependence on voltage	0.1	% full scale/10V	2%/s/10V
losses in the line (leak)	2	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift	
	1.73 mg/m3
	0.11 % value

	ranges		
	min	max	value at calib
flow	18	30	24 l/hr
pressure	99.00	99.00	99.00 kPa
temp	283.00	284.00	283.00 K
NH3 range	0	20	0 mg/m3
N2O range	0	40	0 mg/m3
CH4 range	0	57	0 mg/m3
CO2 range	8	10	0 %vol
H2O range	3.00	3.00	0 %vol
Instrument Voltage Rating	110 V		
Voltage	104.5	115.5	110 V

Measurement performance related to stationary conditions					
Performance characteristic	Uncertainty	Value of uncertainty quantity			
		for mean	for span	use rep at span	
Standard deviation of repeatability at zero	U <sub>r0</sub>				
Standard deviation of repeatability at span level	U <sub>rs</sub>			0.10	
Lack of fit	U <sub>fit</sub>			3.70	
Drift	U <sub>odr</sub>			1.00	
volume or pressure flow dependence	U <sub>spres</sub>			0.11	
atmospheric pressure dependence	U <sub>apres</sub>			0.00	
ambient temperature dependence	U <sub>tamp</sub>			0.00	
NH3 (20 mg/m3)	U <sub>interf</sub>			0.17	
CO2 (15%)	U <sub>interf</sub>			0.18	
H2O (30%)	U <sub>interf</sub>			0.05	
N2O (40 mg/m3)	U <sub>interf</sub>			0.17	
CH4 (57 mg/m3)	U <sub>interf</sub>			0.17	
Dependence on voltage	U <sub>volt</sub>			0.03	
losses in the line (leak)	U <sub>leak</sub>			0.74	
Uncertainty of calibration gas	U <sub>calib</sub>			0.74	
Uncertainty in factor	uf			0.00	

Use largest negative or positive interferent effect		
NH3	0	0.17
CO2	0	0.18
H2O	0	0.05
N2O	0	0.17
CH4	0	0.17
	0	0.75
Interference uncertainty		0.75

Measurement uncertainty	Result	mg/m <sup>3</sup>
Combined uncertainty		4.04
Expanded uncertainty	k = 2	8.09
Uncertainty corrected to std conds		8.34
Expanded uncertainty	expressed with a level of confidence of 95%	#VALUE! % ELV
Expanded uncertainty	expressed with a level of confidence of 95%	8.34 mg.m <sup>-3</sup> at ELV
	U <sub>95</sub> ≤ U <sub>req</sub>	#VALUE!

Verified